



AATCC

ONE DAVIS DRIVE, P O BOX 12215
RESEARCH TRIANGLE PARK, NORTH CAROLINA 27709-2215 USA
TELEPHONE +1-919-549-8141 FAX +1-919-549-8933 <http://www.aatcc.org>

CERTIFICATE OF CONFORMITY

This is to certify that the **AATCC Gray Scale for Color Change, Item # 08359**, supplied to your order has been evaluated and found to be suitable for use in any AATCC Test Method for which it is mentioned. The color chips on this scale fall within the Delta E tolerance ranges listed below and as specified in AATCC Evaluation Procedure 1, Gray Scale for Color Change: Lot # 0211XMUN.

Colorfastness Grade	Total Color Difference CIELAB Units	Tolerance for Working Standards CIELAB Units
5	0.0	+ 0.2
4-5	0.8	± 0.2
4	1.7	± 0.3
3-4	2.5	± 0.3
3	3.4	± 0.4
2-3	4.8	± 0.5
2	6.8	± 0.6
1-2	9.6	± 0.7
1	13.6	± 1.0

Signed:

Date: January 1, 2012

Printed Name: Aaron P. Creech
AATCC Technical Associate