



漢達亞洲有限公司

HCL Asia Ltd.

香港沙田火炭山尾街43-47號環球工業中心6樓14室

Unit 14, 6/F., Worldwide Industrial Centre, 43-47 Shan Mei Street, Fotan, Shatin, Hong Kong

www.hcl-asia.com.hk

T: +852 27420018

F: +852 27420053

info@hcl-asia.com.hk

Textile Moisture Meter

衣服織物濕度測量儀



Model: TEM-I (Aqua-Boy)



50299M

Model: DHT-1



50422M

Calibration Electrode

Using a fixed calibration value to check the accuracy of the display at one test point. The probe is connected to the display which has to show the factory set preset value. The electrode is supplied with a calibration certificate.

校準電極探頭

以一個固定校準值來檢查顯示的精確度是否在預設的測試點範圍內。所述探頭被連接到儀器主機並顯示出廠的預設值。該電極探頭附有一張校準證書。