

INSPECTION CERTIFICATE
検査成績表

Model : 型式	FB-30LB	Serial No. : 製造番号	508623
Capacity : 容量	30lbf	×	0.25lbf
Accuracy 精度	+/- 0.3	%	OF FULL SCALE
Conditions : 測定環境	(Temperature : 温度 23 °C	Humidity : 湿度 52 %)	引張
Date of inspection : 検査年月日	2024/07/22	Tested by : 検査	大木 章睦

Ref. standard 標準	Reading 実測値	Ref. standard 標準	Reading 実測値	Ref. standard 標準	Reading 実測値
0.00	0.00				
5.00	5.00				
10.00	10.00				
15.00	15.00				
20.00	20.00				
25.00	25.00				
30.00	29.97				

NOT FOR PUBLIC RELEASE

We certify that this instrument has been tested according to company procedure.

本製品は、当社社内規格に基づき諸検査を行いました。



Fail
不合格

